

December 26, 2023

Rion Co., Ltd.

### Important Notice

Rion Sound Level Meter NL-43, NL-53, and NL-63

Octave, 1/3 octave Real-time Analysis Program NX-43RT and NX-63RT

Thank you very much for your continued use of our products.

We kindly ask for those who plan to install the octave, 1/3 octave real-time analysis program NX-43RT on the NL-43 or NL-53, or the NX-63RT on the NL-63 to update your sound level meter to the latest firmware version, as stated in the instruction manual of the NX-43RT or NX-63RT.

The program card NX-43RT and NX-63RT stores an upgrade file for the sound level meter at the time of manufacture, and when you insert it into the sound level meter and turn on the power, you can check the version, and if it is an old version, the version will be automatically updated.

The following precautions do not apply to normal usage.

However, if applicable, please be sure to update your sound level meter by following the steps below.

#### [Precautions]

If the version upgrade file (name: update.rup), saved on the program card of NX-43RT or NX-63RT is deleted, and if you plan to install the NX-43RT or NX-63RT to Ver.1.00 sound level meter, please contact the dealer where you purchased our product and receive the latest version of the firmware to install it on your sound level meter.

#### [Inquiries]

Rion Co., Ltd.

S&V Measuring Instruments International Sales

TEL: +81 42-359-7888